

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Specification
8	BME/A2K/(1189)/2023/HTD(A20)	<p><b>ONE OFF PREVENTIVE MAINTENANCE FOR ONE (1) UNIT OF AUTOMATED DNA EXTRACTION SYSTEM, BRAND: QIAGEN, MODEL: QIACUBE CONNECT, S/N: 01794, BME29419 AT FORENSIC BIOLOGY SECTION, DEPARTMENT SCIENTIFIC SERVICES</b></p> <p><u>PLACE OF SUBMISSION:</u>            QUOTATION BOX (GROUND FLOOR)            MINISTRY OF HEALTH            COMMONWEALTH DRIVE            BANDAR SERI BEGAWAN, BB 3910            NEGARA BRUNEI DARUSSALAM</p>	29/11/2023	23/12/2023	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	ZAENUL 'ARIFIEN BIN DATO PADUKA HAJI MOHD ROSLI  JURUTERA PERUBATAN HAYAT  TEL: 2382424 ext. 6122 / 6232

## LAMPIRAN - A1

RUJ: BME/A2K/(1189)/2023/HTD (A20)

21/11/2023

**ONE OFF PREVENTIVE MAINTENANCE FOR ONE (1) UNIT OF AUTOMATED DNA EXTRACTION SYSTEM, BRAND: QIAGEN, MODEL: QIACUBE CONNECT, S/N: 01794, BME29419, AT FORENSIC BIOLOGY SECTION, DEPARTMENT SCIENTIFIC SERVICES**

### SCOPE OF WORK

One-off preventive maintenance for one (1) unit of Automated DNA Extraction system, Brand: Qiagen Model: Qiacube Connect, S/N: 01794, BME29419, at Forensic Biology Section, Department of Scientific Services. Scope of work inclusive but not limited to:

- Scope of work on the Automated DNA Extraction system to follow Qiagen preventive maintenance manual/recommendation specific for this system which include the following where necessary:

- (i) inspection
- (ii) cleaning
- (iii) installation of PM kits, consumables and spare parts listed in spareparts list (Lampiran A2)
- (iv) alignment
- (v) calibration
- (vi) any other related preventive maintenance works required.

- Functional check post-maintenance.

- Apply preventive maintenance sticker/tag after successful service.

- Disposal of any used spare parts/pm kits, unless otherwise instructed.

- Work on the IC system is to be carried out by engineer(s) authorised in providing maintenance to Qiagen Automated DNA Extraction system.

- Vendor(s) to submit maintenance checklist specific for Qiagen Automated DNA Extraction system along with other necessary documents to join this tender.

- Upon approval, the earliest the preventive maintenance work can be carried out on the equipment is one month before the next PM due date, and not anytime earlier.

## LAMPIRAN - A2

RUJ: BME/A2K/(1189)/2023/HTD (A20)

ONE OFF PREVENTIVE MAINTENANCE FOR ONE (1) UNIT OF AUTOMATED DNA EXTRACTION SYSTEM, BRAND: QIAGEN, MODEL: QIACUBE CONNECT, S/N: 01794, BME29419, AT FORENSIC BIOLOGY SECTION, DEPARTMENT SCIENTIFIC SERVICES

NO	SPAREPART	PART NUMBER	QUANTITY
1	PPM	9245212 / 9236451	1