

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Focal Person
3	BME/A2K/(1195)/2023/HTD(A20)	<p><b>ONE OFF PREVENTIVE MAINTENANCE SERVICE FOR ONE (1) UNIT OF MICROWAVE DIGESTER, BRAND: MILESTONE, MODEL: START D, S/N: 131863, BME19895 AT FOOD CHEMISTRY SECTION, DEPARTMENT OF SCIENTIFIC SERVICES</b></p> <p><u>PLACE OF SUBMISSION:</u>            QUOTATION BOX (GROUND FLOOR)            MINISTRY OF HEALTH            COMMONWEALTH DRIVE            BANDAR SERI BEGAWAN, BB 3910            NEGARA BRUNEI DARUSSALAM</p>	02/12/2023	23/12/2023	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	<p>HAJI ZAENUL 'ARIFIEN BIN DATO PADUKA HAJI MOHD ROSLI</p> <p>JURUTERA PERUBATAN HAYAT</p> <p>TEL: 2382424 ext. 6122 / 6232</p>

# LAMPIRAN - A1

RJJ: BME/A2K/(1195)/2023/HTD (A20)

28/11/2023

**ONE-OFF PREVENTIVE MAINTENANCE SERVICE FOR ONE (1) UNIT OF MICROWAVE DIGESTER , BRAND: MILESTONE, MODEL: START D, S/N: 131863, BME19895 AT FOOD CHEMISTRY SECTION, DEPARTMENT OF SCIENTIFIC SERVICES.**

## SCOPE OF WORK

One-off preventive maintenance for one (1) unit of Milestone Start D microwave digester at Food Chemistry Section, Department of Scientific Services, Serasa inclusive of but not limited to:

- Scope to follow preventive maintenance manual/recommendation specific for the equipment, which include the following where necessary:

- (i) inspection
- (ii) cleaning
- (iii) installation of PM kits, consumables and spare parts listed in spareparts list (if any in Lampiran A2)
- (iv) alignment
- (v) calibration
- (vi) any other related preventive maintenance works required.

- Functional check post-maintenance.

- Apply preventive maintenance sticker/tag after successful service.

- Disposal of any used spare parts/pm kits, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

- Upon approval, the earliest the preventive maintenance work can be carried out on the equipment is one month before the next PM due date, and not anytime earlier.