

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Focal Person
1	66/QTN/JPK/2024	<p><b>SUPPLY AND DELIVERY TWENTY (20) UNITS 'PORTABLE EXAMINATION LIGHT' FOR HEALTH CENTRES, DEPARTMENT OF HEALTH SERVICES</b></p> <p><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	26/09/2024	19/10/2024	\$5.00	JABATAN PERKHIDMATAN KESIHATAN, KEMENTERIAN KESIHATAN,	<p>HJH SYAZWANI AZAH BINTI HJ ABU BAKAR</p> <p>JURUTERA PERUBATAN HAYAT</p> <p>TEL: 2381740 ext. 7561 (HTD)</p>

TAWARAN SEBUTHARGA  
 JABATAN PERKHIDMATAN KESIHATAN

Rujukan Tawaran: **66/QTN/JPK/2024**

Nama Tawaran:

**SUPPLY AND DELIVERY TWENTY (20) UNITS 'PORTABLE EXAMINATION LIGHT' FOR HEALTH CENTRES, DEPARTMENT OF HEALTH SERVICES.**

<b>SPECIFICATIONS</b>	
<b>STANDARD FEATURES</b>	<p><b>PORTABLE EXAMINATION LIGHT – 20 UNITS</b></p> <ul style="list-style-type: none"> <li>- Heavy base with low Centre of gravity</li> <li>- Articulating arm between light and base stand</li> <li>- Base with 5 anti-static swivel castors in star formation</li> <li>- At least two castors have been equipped with brakes</li> <li>- Minimum 14cm fixed focus field of view</li> <li>- Minumun illumination of 60,000 lux at 0.5m</li> <li>- LED light source, with minimum LED life more than 50,000 hours.</li> <li>- Floor to light height adjustable to include at least range og 1.1m to 1.75m</li> <li>- Coiled goose neck. Solid construction with high-specification of at least 100cm radius</li> <li>- On/off switch</li> </ul>

• Authorized by:

.....

TAWARAN SEBUTHARGA  
JABATAN PERKHIDMATAN KESIHATAN

---

Rujukan Tawaran: **66/QTN/JPK/2024**

Nama Tawaran:

**SUPPLY AND DELIVERY TWENTY (20) UNITS 'PORTABLE EXAMINATION LIGHT' FOR HEALTH CENTRES, DEPARTMENT OF HEALTH SERVICES.**

SPECIFICATIONS	
STANDARD FEATURES	TWENTY (20) UNITS PORTABLE EXAMINATION LIGHT

**Please state:**

- Brand/ Model
- Accessories
- Country of origin
- International Safety Standard
- Year Manufactured
- Price per unit
- Total price
- Delivery period
- Price Validity
- Sample & Brochure
- Warranty
- 

• Authorized by:

.....

SPECIFICATIONS  
DISTRIBUTIONS:

FORM A

	<b>Health Centre / Clinic</b>	<b>Units</b>
1	Pusat Kesihatan Berakas	10
2	PAPHMWHB – Gadong HC	2
3	Pusat Kesihatan Muara	2
4	Pusat Kesihatan Pekan Tutong	1
5	Pusat Kesihatan Kuala Belait	1
6	Pusat Kesihatan Sungai Liang	1
7	Pusat Kesihatan Berakas - MCH	1
8	Pusat Kesihatan Telisai - MCH	1
9	Tambahan	1
	<b>Total</b>	<b>20</b>

**TAWARAN SEBUTHARGA  
JABATAN PERKHIDMATAN KESIHATAN**

Rujukan Tawaran: **66/QTN/JPK/2024**

Nama Tawaran:

**SUPPLY AND DELIVERY TWENTY (20) UNITS 'PORTABLE EXAMINATION LIGHT' FOR HEALTH CENTRES, DEPARTMENT OF HEALTH SERVICES.**

SPECIFICATIONS		YES	NO	IF 'NO' PLEASE STATE ALTERNATIVE OFFER
<b>STANDARD FEATURES</b>	Heavy base with low Centre of gravity			
	Articulating arm between light and base stand			
	Base with 5 anti-static swivel castors in star formation			
	At least two castors have been equipped with brakes			
	Minimum 14cm fixed focus field of view			
	Minumun illumination of 60,000 lux at 0.5m			
	LED light source, with minimum LED life more than 50,000 hours			
	Floor to light height adjustable to include at least range og 1.1m to 1.75m			
	Coiled goose neck. Solid construction with high-specification of at least 100cm radius			
	On/off switch			

<b>PROCUREMENT REQUIREMENT</b>	
Brand/ Model	
Accessories	
Country of origin	
International Safety Standard	
Dimensions	
Weight	
Materials	
Year Manufactured	
Price per unit	
Total price	
Delivery period	
Price Validity	
Sample & Brochure	
Warranty	