

Rujukan Kami: **(46) MOH/HQ/P/IKLAN-SH/2024**

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Focal Person
7	(33) PSD/QTN/2024 – (SSBH)	<p>SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS TARGET CONTROLLED INFUSION (TCI) SYRINGE PUMP FOR OPERATING THEATER, SURI SERI BEGAWAN (SSB) HOSPITAL, KUALA BELAIT(CLUSTERING)</p> <p>PLACE OF SUBMISSION: QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	26/03/2024	20/04/2024	\$5.00	HOSPITAL SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT.	<p>Danil bin Osman Pegawai Jururawat Bilik Bedah Hospital Suri Seri Begawan</p> <p>Tel : 3335331 Ext : 3146</p>

ITEM(S) SPECIFICATIONS FOR ADVERTISEMENT (ABOVE \$2000)

QUOTATION/TENDER REFERENCE NO:	
QUOTATION/TENDER NAME:	SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS TARGET CONTROLLED INFUSION (TCI) SYRINGE PUMP FOR OPERATING THEATER, SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT (CLUSTERING).

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY OFFERED	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS (B\$)
1	TARGET CONTROLLED INFUSION (TCI) SYRINGE PUMP	UNIT	2						
TOTAL PRICE (B\$)									

NO	<u>TERMS AND CONDITIONS</u>	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	QUOTATION/TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF QUOTATION/TENDER .	
4	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION/TENDER .	
5	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery . Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
6	Brochures / catalogues should be submitted / attached with quotation/tender document.	
7	Samples should be submitted together with quotation/ tender or within fourteen (14 days) of the quotation/tender closing date (if applicable).	
8	DELIVERY PERIOD: Not later than 8 weeks	(Yes / No) (If No, please specify)
9	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

FOR QUOTATION ONLY

FOR QUOTATION ONLY		TERMS AND CONDITIONS	
a.	Tenderer must be registered with the Ministry of Health.	<i>Acknowledgement:</i>	Company's Official Stamp
b.	Please fill in the QUOTATION FORM completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF QUOTATION .	<i>Company Ref. No.:</i> I hereby certify the above quote to be correct. Signature:	
c.	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION .	Name:	
d.	Please do not use TIPPEX for amendment.	Designation: Date:	

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2)
 UNITS TARGET CONTROLLED INFUSION (TCI) SYRINGE PUMP FOR OPERATING
 THEATER, SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT**

	USER REQUIREMENT
STANDARD FEATURES:	<ul style="list-style-type: none"> Must be intended for use in a surgical setting/environment Intended to allow the automation of dosing of intravenous drugs during surgery monitored by the anesthetist
	TECHNICAL SPECIFICATIONS:
	<ul style="list-style-type: none"> Must be TIVA/TCI capable
	<ul style="list-style-type: none"> Must be compatible with Terumo brand Syringes currently being utilized by the Ministry of Health, Brunei
	<ul style="list-style-type: none"> Infusion/Delivery Rate Minimum: 0.01 mL/Hr
	<ul style="list-style-type: none"> Infusion/Delivery Rate Maximum: 1800 mL/Hr
	<ul style="list-style-type: none"> Infusion/Delivery Rate Accuracy: Less than 5%
	<ul style="list-style-type: none"> Must be able to infuse the following drugs: Propofol, Remifentanyl
	<ul style="list-style-type: none"> TCI models expected to be included to be used with Propofol but are not limited to: Marsh, Schnider, Eleveld
	<ul style="list-style-type: none"> TCI models expected to be included to be used with Remifentanyl but are not limited to: Minto
	<ul style="list-style-type: none"> Must have a Dose Error Reduction System (Software)
	<ul style="list-style-type: none"> Must have alarms to detect the following states but are not limited to: High Pressure/Occlusion, Empty Syringe, Infusion near end, Low Battery, Depleted Battery
	POWER:
	<ul style="list-style-type: none"> Must be compatible with 220-240 V AC, 50-60 Hz mains power supply
	<ul style="list-style-type: none"> Must have an integrated rechargeable battery with 8 hours minimum battery life
	MISCELLANEOUS:
	<ul style="list-style-type: none"> Any essential accessories required for the day to day operation of the unit and not mentioned above such as cables or pole clamps must be supplied for each unit

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2)
 UNITS TARGET CONTROLLED INFUSION (TCI) SYRINGE PUMP FOR OPERATING
 THEATER, SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT**

SECTION II - PROCUREMENT REQUIREMENT	
MODEL AND BRAND	<ul style="list-style-type: none"> State Model and Brand of Equipment.
COUNTRY OF ORIGIN	<ul style="list-style-type: none"> State Country of Origin.
PRICE (CIF) B\$	<ul style="list-style-type: none"> State Unit and Total Price Offered.
WHERE MARKETING	<ul style="list-style-type: none"> Specify in which other countries the equipment marketed.
YEAR MANUFACTURED	<ul style="list-style-type: none"> State year of equipment manufactured.
WARRANTY	<ul style="list-style-type: none"> State number of years.
DELIVERY TIME	<ul style="list-style-type: none"> Not more than 90 days.
PRICE VALIDITY	<ul style="list-style-type: none"> State number of weeks.
TECHNICAL SPECIFICATIONS/REQUIREMENTS	
POWER REQUIREMENT:	<ul style="list-style-type: none"> Mains Supply 240VAC 50 Hz.
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	<ul style="list-style-type: none"> State equipment ambient operating temperature range.
INTERNATIONAL SAFETY STANDARD:	<ul style="list-style-type: none"> The equipment must comply to either one of the five international safety standard, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese JIS Standard.
TECHNICAL SUPPORT:	<ul style="list-style-type: none"> State number of locally stationed engineers/technicians and the nearest overseas support.
DIMENSIONS:	<ul style="list-style-type: none"> State overall dimensions of equipment.
WEIGHT:	<ul style="list-style-type: none"> State net weight of the equipment.
EQUIPMENT WHOLE LIFE TIME SUPPORT:	<ul style="list-style-type: none"> Supplier must guarantee spare parts are available at least for 10 years after the installation of the equipment.
BROCHURE:	<ul style="list-style-type: none"> Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.
USER AND SERVICE MANUALS:	<ul style="list-style-type: none"> Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.

