

Rujukan Kami: **(54) MOH/HQ/P/IKLAN-SH/2024**

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Focal Person
1	DSS/FS/PROC/B50K/2024-2025/002(ILD)	<p>TO SUPPLY AND DELIVER VOLUMETRIC FLASKS OF MULTIPLE VOLUMES FOR ILLICIT DRUGS SECTION AND TOXICOLOGY SECTION OF FORENSIC SCIENCE DIVISION, DEPARTMENT OF SCIENTIFIC SERVICES, MNISTRY OF HEALTH (CLUSTERING)</p> <p><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	22/04/2024	04/05/2024	\$5.00	JABATAN PERKHIDMATAN SAINTIFIK, KEMENTERIAN KESIHATAN.	Pg Mohd Helmy bin Pg Hj Rabaha ILLICIT Drugs Section Tel : 2384213

ITEM(S) SPECIFICATIONS FOR ADVERTISEMENT

TENDER/QUOTATION NAME	TO SUPPLY AND DELIVER VOLUMETRIC FLASKS OF MULTIPLE VOLUMES FOR ILLICIT DRUGS SECTION AND TOXICOLOGY SECTION OF FORENSIC SCIENCE DIVISION , DEPARTMENT OF SCIENTIFIC SERVICES, MINISTRY OF HEALTH (CLUSTERING).
------------------------------	---

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
1	5 mL Volumetric Flask Class A	<ul style="list-style-type: none"> • Complies with ISO 1042 and DIN 12664. • Manufactured from borosilicate glass 3.3. • Calibrated to the tolerances required by the United States Pharmacopoeia (USP). • Each flask must be individually supplied with a calibration certificate complying to the requirements of ISO 17025. • Calibration must be by ISO 17025 calibration laboratory. • Each flask must be labelled with an individual serial number and date of manufacture for traceability purposes. • White enamel inscriptions. • With Polyethylene stopper. 	3151/04 UKAS	1 Pc	35 Pcs					

Procurement Ref No : DSS/FS/PROC/B50K/2024-2025/002(ILD)
 Section Ref No : ILD/PROC/B50K/2404

2	10 mL Volumetric Flask Class A	<ul style="list-style-type: none"> Complies with ISO 1042 and DIN 12664. Manufactured from borosilicate glass 3.3. Calibrated to the tolerances required by the United States Pharmacopoeia (USP). Each flask must be individually supplied with a calibration certificate complying to the requirements of ISO 17025. Calibration must be by ISO 17025 calibration laboratory. Each flask must be labelled with an individual serial number and date of manufacture for traceability purposes. White enamel inscriptions. With Polyethylene stopper. 	3151/06 UKAS	1 Pc	40 Pcs					
3	20 mL Volumetric Flask Class A	<ul style="list-style-type: none"> Complies with ISO 1042 and DIN 12664. Manufactured from borosilicate glass 3.3. Calibrated to the tolerances required by the United States Pharmacopoeia (USP). Each flask must be individually supplied with a calibration certificate complying to the requirements of ISO 17025. Calibration must be by ISO 17025 calibration laboratory. Each flask must be labelled with an individual serial number and date of manufacture for traceability purposes. White enamel inscriptions. With Polyethylene stopper. 	3151/08 UKAS	1 Pc	60 Pcs					

4	25 mL Volumetric Flask Class A	<ul style="list-style-type: none"> Complies with ISO 1042 and DIN 12664. Manufactured from borosilicate glass 3.3. Calibrated to the tolerances required by the United States Pharmacopoeia (USP). Each flask must be individually supplied with a calibration certificate complying to the requirements of ISO 17025. Calibration must be by ISO 17025 calibration laboratory. Each flask must be labelled with an individual serial number and date of manufacture for traceability purposes. White enamel inscriptions. With Polyethylene stopper. 	3151/10 UKAS	1 Pc	250 Pcs					
5	50 mL Volumetric Flask Class A	<ul style="list-style-type: none"> Complies with ISO 1042 and DIN 12664. Manufactured from borosilicate glass 3.3. Calibrated to the tolerances required by the United States Pharmacopoeia (USP). Each flask must be individually supplied with a calibration certificate complying to the requirements of ISO 17025. Calibration must be by ISO 17025 calibration laboratory. Each flask must be labelled with an individual serial number and date of manufacture for traceability purposes. White enamel inscriptions. With Polyethylene stopper.inscriptions. 	3151/12 UKAS	1 Pc	60 Pcs					

6	100 mL Volumetric Flask Class A	<ul style="list-style-type: none"> Complies with ISO 1042 and DIN 12664. Manufactured from borosilicate glass 3.3. Calibrated to the tolerances required by the United States Pharmacopoeia (USP). Each flask must be individually supplied with a calibration certificate complying to the requirements of ISO 17025. Calibration must be by ISO 17025 calibration laboratory. Each flask must be labelled with an individual serial number and date of manufacture for traceability purposes. White enamel inscriptions. With Polyethylene stopper. 	3151/14 UKAS	1 Pc	50 Pcs				
7	250 mL Volumetric Flask Class A	<ul style="list-style-type: none"> Complies with ISO 1042 and DIN 12664. Manufactured from borosilicate glass 3.3. Calibrated to the tolerances required by the United States Pharmacopoeia (USP). Each flask must be individually supplied with a calibration certificate complying to the requirements of ISO 17025. Calibration must be by ISO 17025 calibration laboratory. Each flask must be labelled with an individual serial number and date of manufacture for traceability purposes. White enamel inscriptions. With Polyethylene stopper. 	3151/18 UKAS	1 Pc	30 Pcs				
8	Warranty	<ul style="list-style-type: none"> 1 year warranty or better. 	-	-	-				
9	Delivery	<ul style="list-style-type: none"> Mode of delivery must be air freight. 	-	-	-				

Procurement Ref No : DSS/FS/PROC/B50K/2024-2025/002(ILD)
Section Ref No : ILD/PROC/B50K/2404

NOTE	VENDOR'S OFFER
<i>Delivery Period Upon the Issuance of PO</i>	
<i>Price Validity must not be less than 90 days</i>	

Bahagian/Unit	ILLCIT DRUGS SECTION		
Bil. Rujukan Bahagian/Unit*:	ILD/PROC/B50K/2404		
Pegawai di rujuk	Nama	PG MOHD HELMY BIN PG HJ RABAHA	
	E-mail	Helmy.Rabaha @ MOH.GOV.BN	
	Tel.No.	2384213	Fax No. 2381946
UNTUK KEGUNAAN BAHAGIAN PENTADBIRAN SAHAJA			
No. Rujukan PROCUREMENT			
Tarikh Menghantar Iklan :			

Procurement Ref No : DSS/FS/PROC/B50K/2024-2025/002(ILD)
Section Ref No : ILDC/PROC/B50K/2404