

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
2	BME/A2K/(970)/2022/HTD (A17)	<p><b>TO SERVICE ONE (1) UNIT OF INNOGENETIC AUTO-LIA48 AUTOMATED HIV ANALYZER, BME5363 AT VIROLOGY LAB, SUMBILING, MINISTRY OF HEALTH</b></p> <p><u>PLACE OF SUBMISSION:</u>            QUOTATION BOX (GROUND FLOOR)            MINISTRY OF HEALTH COMMONWEALTH DRIVE            BANDAR SERI BEGAWAN, BB 3910            NEGARA BRUNEI DARUSSALAM</p>	05/01/2023	21/01/2023	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	<p>DK HANI MAHYANA BINTI PG HAJI DAMIT</p> <p>JURUTERA PERUBATAN HAYAT</p> <p>TEL: 2242424 ext. 8693 (RIPASH)</p>

# LAMPIRAN - A1

RUJ: BME/A2K/(970)/2022/HTD (A17)

21/12/2022

TO SERVICE ONE (1) UNIT OF INNOGENETICS AUTO-LIA48 AUTOMATED HIV ANALYZER, BME5363 AT VIROLOGY LAB, SUMBILING, MINISTRY OF HEALTH

## SCOPE OF WORK

### EQUIPMENT DETAILS:

DESCRIPTION : AUTOMATED HIV ANALYZER

BRAND : INNOGENETICS

MODEL : AUTO-LIA-48

BME NO: BME5363

S/N : 501000002

### SCOPE OF WORK:

VERIFY SYSTEM USING FLOPPY DISKETTE OF THE SYSTEM

CHECK AND CLEAN OPTIONAL PRINTER

CHECK AND CLEAN DISPLAY MONITOR MODULE

CLEAN AND GREASE GUIDING BARS

CHECK AND CLEAN PUMP HEAD

CLEAN VERIFIER MODULE TEMPERATURE

CLEAN AND PERFORM PUMP TEST

CHANGE CALIBRATION TUBING

CLEAN LEVEL SENSOR PIN IN TUBING

CHECK & CLEAN FRONT COOLING FAN

CHECK AC/DC POWER SUPPLY VOLTAGE

REPLACE ALL DISPENSING TUBES

PERFORM TUBE VOLUME ADJUSTMENT

PERFORM BATTERY MEASUREMENT

PERFORM PUMP CALIBRATION

PERFORM CHANNEL ADJUSTMENT

PERFORM PUMP VOLUME RECALIBRATION

GENERAL CLEANING ON THE INSTRUMENT

### PARTS TO BE REPLACED (IF NECESSARY) :

DISPENSING TUBES (INCLUDED IN PM)

CALIBRATION TUBING (INCLUDED IN PM)

## LAMPIRAN - A2

RUJ: BME/A2K/(970)/2022/HTD (A17)

TO SERVICE ONE (1) UNIT OF INNOGENETICS AUTO-LIA48 AUTOMATED HIV ANALYZER, BME5363 AT VIROLOGY LAB, SUMBILING, MINISTRY OF HEALTH

NO	SPAREPART	PART NUMBER	QUANTITY
----	-----------	-------------	----------