

| BIL | Quotation Reference | Description  | Advertisement Date | Closing Date (Not Later Than 2.00PM) | Quotation Fee | Requesting Department                                  | Focal Person   |
|-----|---------------------|--|--------------------|--------------------------------------|---------------|--|--|
| 2   | 38/QTN/JPK/2022     | <p>SUPPLY, DELIVERY INSTALLATION, TESTING AND COMMISSIONING, SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH</p> <p>PLACE OF SUBMISSION:<br/>           QUOTATION BOX (GROUND FLOOR)<br/>           MINISTRY OF HEALTH<br/>           COMMONWEALTH DRIVE<br/>           BANDAR SERI BEGAWAN, BB 3910<br/>           NEGARA BRUNEI DARUSSALAM</p> | 16/11/2022         | 03/12/2022                           | \$5.00        | JABATAN PERKHIDMATAN KESIHATAN, KEMENTERIAN KESIHATAN. | <p>Haji Amir Hisma Bin Masri</p> <p>JURUTERA PERUBATAN HAYAT</p> <p>TEL: 2381640</p> |

**UNIT TAWARAN  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,  
SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW  
BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES  
DEPARTMENT, MINISTRY OF HEALTH**

|                           | <b>SECTION I - USER REQUIREMENT</b>  |
|---------------------------|--|
| <b>STANDARD FEATURES:</b> |  |
|                           | <ul style="list-style-type: none"> <li>• Professional Mercury-free Manual Sphygmomanometers design for clinical use in a professional setting including a physician's office, hospital or other point of care environment.</li> </ul>  |
|                           | <ul style="list-style-type: none"> <li>• Desk model unit with scratch free base and preferably with carrying handle.</li> </ul>  |
|                           | <ul style="list-style-type: none"> <li>• Large and easy-to-read display</li> </ul>   |
|                           | <ul style="list-style-type: none"> <li>• Latex free inflation system with coiled tubing</li> </ul>   |
|                           | <ul style="list-style-type: none"> <li>• Range : approx.. 0 – 300 mmHg</li> </ul>  |
|                           | <ul style="list-style-type: none"> <li>• Includes two Arm Cuff sizes (with tubing connectors) that can easily be swapped (depending on patient):               <ul style="list-style-type: none"> <li>➢ One (1) Standard adult arm cuff &amp; bladder set, double tube with bulb and valve (approx. 24 - 34 cm arm circumference, cuff width approx. 15cm).</li> <li>➢ One (1) large adult arm cuff &amp; bladder set, double tube with bulb and valve (approx. 28-42 cm arm circumference, cuff width approx. 15cm).</li> </ul> </li> </ul> |
|                           | <ul style="list-style-type: none"> <li>• Supplied with durable carrying casing</li> </ul>  |

**UNIT TAWARAN  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,  
SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW  
BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES  
DEPARTMENT, MINISTRY OF HEALTH**

|                           | <b>SECTION II - PROCUREMENT REQUIREMENT</b>  |
|---------------------------|--|
| <b>MODEL AND BRAND:</b>   | <ul style="list-style-type: none"> <li>• State Model and Brand of Equipment.</li> </ul>                      |
| <b>COUNTRY OF ORIGIN:</b> | <ul style="list-style-type: none"> <li>• State Country of Origin.</li> </ul>                                 |
| <b>PRICE (CIF) B\$:</b>   | <ul style="list-style-type: none"> <li>• State Unit and Total Price Offered.</li> </ul>                      |
| <b>WHERE MARKETED:</b>    | <ul style="list-style-type: none"> <li>• Specify in which other countries the equipment marketed.</li> </ul> |
| <b>YEAR MANUFACTURED:</b> | <ul style="list-style-type: none"> <li>• State year of equipment manufactured.</li> </ul>                    |
| <b>WARRANTY:</b>          | <ul style="list-style-type: none"> <li>• State number of years.</li> </ul>                                   |
| <b>DELIVERY TIME:</b>     | <ul style="list-style-type: none"> <li>• State number of weeks.</li> </ul>                                   |
| <b>PRICE VALIDITY:</b>    | <ul style="list-style-type: none"> <li>• State number of weeks.</li> </ul>                                   |

**UNIT TAWARAN  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,  
SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW  
BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES  
DEPARTMENT, MINISTRY OF HEALTH**

|   | <b>SECTION III - TECHNICAL REQUIREMENT</b>   |
|---|--|
| <b>MAINS POWER SUPPLY:</b>                            | <ul style="list-style-type: none"> <li>• Mains Supply 240VAC 50Hz, Battery Backup.</li> </ul>  |
| <b>EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:</b> | <ul style="list-style-type: none"> <li>• State equipment ambient operating temperature range.</li> </ul>   |
| <b>INTERNATIONAL SAFETY STANDARD:</b>                 | <ul style="list-style-type: none"> <li>• The equipment must comply to either one of the five international safety standard, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese Standard.</li> </ul> |
| <b>TECHNICAL SUPPORT:</b>                             | <ul style="list-style-type: none"> <li>• State number of locally stationed engineers/technicians and the nearest overseas support.</li> </ul>  |
| <b>DIMENSIONS:</b>                                    | <ul style="list-style-type: none"> <li>• State overall dimensions of equipment.</li> </ul>   |
| <b>WEIGHT:</b>  | <ul style="list-style-type: none"> <li>• State nett weight of the equipment.</li> </ul>  |
| <b>EQUIPMENT WHOLE LIFE TIME SUPPORT:</b>             | <ul style="list-style-type: none"> <li>• Number of years the spare parts are available after the installation of the equipment.</li> </ul>   |
| <b>BROUCHER:</b>                                      | <ul style="list-style-type: none"> <li>• Tenderers must submit with their offer, a detailed broucher which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.</li> </ul>                                     |
| <b>USER AND SERVICE MANUALS:</b>                      | <ul style="list-style-type: none"> <li>• Two hard copies and one soft copy [ PDF FORMAT ] to be submitted to BME office during commissioning.</li> </ul>   |
| <b>SPARE-PARTS AND CONSUMABLES LISTINGS:</b>          | <ul style="list-style-type: none"> <li>• Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.</li> </ul>                                  |

**UNIT TAWARAN  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,  
SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW  
BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES  
DEPARTMENT, MINISTRY OF HEALTH**

**UNIT TAWARAN**  
**KEMENTERIAN KESIHATAN**  
**NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH**

|                           | <b>SECTION I - USER REQUIREMENT</b>  | <b>YES</b> | <b>NO</b> | <b>(IF 'NO')<br/>Please State<br/>Alternative Offer</b> |
|---------------------------|--|------------|-----------|---|
| <b>STANDARD FEATURES:</b> |  |            |           |   |
|                           | <ul style="list-style-type: none"> <li>Professional Mercury-free Manual Sphygmomanometers design for clinical use in a professional setting including a physician's office, hospital or other point of care environment.</li> </ul>  |            |           |   |
|                           | <ul style="list-style-type: none"> <li>Desk model unit with scratch free base and preferably with carrying handle.</li> </ul>  |            |           |   |
|                           | <ul style="list-style-type: none"> <li>Large and easy-to-read display</li> </ul>   |            |           |   |
|                           | <ul style="list-style-type: none"> <li>Latex free inflation system with coiled tubing</li> </ul>   |            |           |   |
|                           | <ul style="list-style-type: none"> <li>Range : approx. 0 – 300 mmHg</li> </ul>   |            |           |   |
|                           | <ul style="list-style-type: none"> <li>Includes two Arm Cuff sizes (with tubing connectors) that can easily be swapped (depending on patient):               <ul style="list-style-type: none"> <li>➤ One (1) Standard adult arm cuff &amp; bladder set, double tube with bulb and valve (approx. 24 - 34 cm arm circumference, cuff width approx. 15cm).</li> </ul> </li> </ul> |            |           |   |

**UNIT TAWARAN**  
**KEMENTERIAN KESIHATAN**  
**NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH**

|  |  |  |  |  |
|--|--|--|--|--|
|  | <ul style="list-style-type: none"> <li>➤ One (1) large adult arm cuff &amp; bladder set, double tube with bulb and valve (approx. 28-42 cm arm circumference, cuff width approx. 15cm).</li> </ul> |  |  |  |
|  | <ul style="list-style-type: none"> <li>• Supplied with durable carrying casing</li> </ul>  |  |  |  |

**UNIT TAWARAN**  
**KEMENTERIAN KESIHATAN**  
**NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH**

| SECTION II – PROCUREMENT REQUIREMENT |                              |                               |
|--------------------------------------|------------------------------|-------------------------------|
| <b>MODEL AND BRAND</b>               |                              |                               |
| <b>COUNTRY OF ORIGIN:</b>            |                              |                               |
| <b>PRICE (CIF) B\$</b>               | <b>UNIT PRICE (B\$ CIF):</b> | <b>TOTAL PRICE (B\$ CIF):</b> |
| <b>WHERE MARKETED:</b>               |                              |                               |
| <b>YEAR MANUFACTURED:</b>            |                              |                               |
| <b>WARRANTY:</b>                     |                              |                               |
| <b>DELIVERY TIME:</b>                |                              |                               |
| <b>PRICE VALIDITY:</b>               |                              |                               |
| SECTION III - TECHNICAL REQUIREMENTS |                              |                               |



**UNIT TAWARAN  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH**

|   |  |            |           |
|---|--|------------|-----------|
|   |  |            |           |
| <b>MAINS POWER SUPPLY:</b>                            |  |            |           |
| <b>EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:</b> |  |            |           |
| <b>BATTERY BACKUP:</b>                                |  |            |           |
| <b>INTERNATIONAL SAFETY STANDARD:</b>                 |  |            |           |
| <b>TECHNICAL SUPPORT:</b>                             |  |            |           |
| <b>DIMENSIONS(mm):</b>                                |  |            |           |
| <b>WEIGHT(in Kg):</b>                                 |  |            |           |
| <b>EQUIPMENT WHOLE LIFE TIME SUPPORT:</b>             | Number of years, spare parts are available after the installation of the equipment: _____ years. |            |           |
|   |  | <b>YES</b> | <b>NO</b> |

**UNIT TAWARAN**  
**KEMENTERIAN KESIHATAN**  
**NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, SIXTEEN (16) UNITS OF NON-MERCURY MANUAL SPHYGMOMANOMETERS FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH**

|   |   |  |  |
|---|---|--|--|
| <b>BROUCHER</b>                             | Detailed broucher submitted?  |  |  |
| <b>USER MANUALS</b>                         | Tenderers must supply during commissioning, three sets of user manuals, one set must be in the form of soft copy. ( PDF Format )  |  |  |
| <b>SERVICE MANUALS</b>                      | Tenderers must supply during commissioning, three sets of service manuals, one set must be in the form of soft copy. ( PDF Format ).  |  |  |
| <b>SPARE-PARTS AND CONSUMABLES LISTINGS</b> | <p>Tenderers must supply during commissioning, a comprehensive list of equipment spare-parts and consumables in the form of soft copy (e.g: Microsoft EXCEL).</p> <p>The listing must have the following details;</p> <ul style="list-style-type: none"> <li>• Part Number</li> <li>• Part Description</li> </ul> |  |  |