

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
5	41/QTN/JPK/2022	<p>SUPPLY, DELIVERY INSTALLATION, TESTING AND COMMISSIONING, FIVE (5) UNITS OF ADULT HEIGHT & WEIGHING SCALE FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH</p> <p><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	16/11/2022	03/12/2022	\$5.00	JABATAN PERKHIDMATAN KESIHATAN, KEMENTERIAN KESIHATAN.	HAJI AMIR HISMA BIN MASRI JURUTERA PERUBATAN HAYAT TEL: 2381640

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING FIVE (5) UNITS OF
 ADULT HEIGHT & WEIGHING SCALE FOR NEW BLOCK SILVER JUBILEE SENGKURONG
 HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH**

	USER REQUIREMENT
STANDARD FEATURES:	
	<ul style="list-style-type: none"> • Standing weighing scale with large and easy to read LCD display.
	<ul style="list-style-type: none"> • With HOLD functions and Auto Shut-Off.
	<ul style="list-style-type: none"> • Fitted with the telescopic height measuring rod.
	<ul style="list-style-type: none"> • Standing weighing scale with handrails
	<ul style="list-style-type: none"> • Operated on AC Adapter AND Alkaline Batteries.
	<ul style="list-style-type: none"> • Weight capacity must not be less than 450 lbs / 200 kgs.
	<ul style="list-style-type: none"> • Supplied with 240VAC Mains Adapter.

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING FIVE (5) UNITS OF
ADULT HEIGHT & WEIGHING SCALE FOR NEW BLOCK SILVER JUBILEE SENGKURONG
HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH**

SECTION II - PROCUREMENT REQUIREMENT	
MODEL AND BRAND	<ul style="list-style-type: none"> State Model and Brand of Equipment.
COUNTRY OF ORIGIN	<ul style="list-style-type: none"> State Country of Origin.
PRICE (CIF) B\$	<ul style="list-style-type: none"> State Unit and Total Price Offered.
WHERE MARKETED	<ul style="list-style-type: none"> Specify in which other countries the equipment marketed.
YEAR MANUFACTURED	<ul style="list-style-type: none"> State year of equipment manufactured.
WARRANTY	<ul style="list-style-type: none"> State number of years.
DELIVERY TIME	<ul style="list-style-type: none"> Not more than 90 days.
PRICE VALIDITY	<ul style="list-style-type: none"> State number of weeks.
TECHNICAL SPECIFICATIONS/REQUIREMENTS	
POWER REQUIREMENT:	<ul style="list-style-type: none"> Mains Supply 240VAC 50 Hz.
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	<ul style="list-style-type: none"> State equipment ambient operating temperature range.
INTERNATIONAL SAFETY STANDARD:	<ul style="list-style-type: none"> The equipment must comply to either one of the five international safety standard, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese JIS Standard.
TECHNICAL SUPPORT:	<ul style="list-style-type: none"> State number of locally stationed engineers/technicians and the nearest overseas support.
DIMENSIONS:	<ul style="list-style-type: none"> State overall dimensions of equipment.
WEIGHT:	<ul style="list-style-type: none"> State net weight of the equipment.
EQUIPMENT WHOLE LIFE TIME SUPPORT:	<ul style="list-style-type: none"> Supplier must guarantee spare parts are available at least for 10 years after the installation of the equipment.
BROCHURE:	<ul style="list-style-type: none"> Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.
USER AND SERVICE MANUALS:	<ul style="list-style-type: none"> Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING FIVE (5) UNITS OF
ADULT HEIGHT & WEIGHING SCALE FOR NEW BLOCK SILVER JUBILEE SENGKURONG
HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH**

SPARE-PARTS AND CONSUMABLES LISTINGS:	<ul style="list-style-type: none"> • Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.
WARRANTY PLANNED PREVENTIVE MAINTENANCE:	<ul style="list-style-type: none"> • Minimum of two times PPM per year during the equipment warranty, six months after the date of commissioning and at the end of warranty period.
TRAINING: ON-SITE:	<ul style="list-style-type: none"> • Tenderers are required to conduct technical training to Biomedical Engineers and Technicians. • Training to be conducted locally, tenderers are required to : <ul style="list-style-type: none"> • Provide training materials, test equipment, demo equipments, etc. • Provide training to two groups of technical staffs. • Provide 2 days (minimum) of training for each group. • Training to be conducted at Ministry of Health

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING FIVE (5) UNITS OF ADULT HEIGHT & WEIGHING SCALE FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH

	USER REQUIREMENTS	<i>Please state your offer</i>
STANDARD FEATURES:		
	<ul style="list-style-type: none"> • Standing weighing scale with large and easy to read LCD display. 	
	<ul style="list-style-type: none"> • With HOLD functions and Auto Shut-Off. 	
	<ul style="list-style-type: none"> • Fitted with the telescopic height measuring rod. 	
	<ul style="list-style-type: none"> • Standing weighing scale with handrails 	
	<ul style="list-style-type: none"> • Operated on AC Adapter AND Alkaline Batteries. 	
	<ul style="list-style-type: none"> • Weight capacity must not be less than 450 lbs / 200 kgs. 	
	<ul style="list-style-type: none"> • Supplied with 240VAC Mains Adapter. 	

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING FIVE (5) UNITS OF ADULT HEIGHT & WEIGHING SCALE FOR NEW
BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH**

SECTION II - PROCUREMENT REQUIREMENT			
BRAND AND MODEL OF EQUIPMENT:			
COUNTRY OF ORIGIN:			
PRICE (CIF) B\$	<table border="1" style="width: 100%;"> <tr> <td style="width: 60%;">UNIT PRICE (B\$ CIF):</td> <td>TOTAL PRICE (B\$ CIF):</td> </tr> </table>	UNIT PRICE (B\$ CIF):	TOTAL PRICE (B\$ CIF):
UNIT PRICE (B\$ CIF):	TOTAL PRICE (B\$ CIF):		
WHERE MARKETED:			
YEAR MANUFACTURED:			
WARRANTY:			
DELIVERY TIME:			
PRICE VALIDITY:			

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING FIVE (5) UNITS OF ADULT HEIGHT & WEIGHING SCALE FOR NEW
BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH**

SECTION III - TECHNICAL REQUIREMENTS	
MAINS POWER SUPPLY:	
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	
BATTERY BACKUP:	
INTERNATIONAL SAFETY STANDARD:	
TECHNICAL SUPPORT:	
DIMENSIONS (mm):	
WEIGHT (in kg):	
EQUIPMENT WHOLE LIFE TIME SUPPORT:	Number of years, spare parts are available after the installation of the equipment: _____ years

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING FIVE (5) UNITS OF ADULT HEIGHT & WEIGHING SCALE FOR NEW
BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH**

		YES	NO
BROCHURE:	Detailed brochure submitted?		
USER MANUALS:	Tenderers must supply during commissioning, three sets of user manuals, one set must be in the form of soft copy. (PDF Format)		
SERVICE MANUALS:	Tenderers must supply during commissioning, three sets of service manuals, one set must be in the form of soft copy. (PDF Format).		
SPARE-PARTS AND CONSUMABLES LISTINGS:	Tenderers must supply during commissioning, a comprehensive list of equipment spare-parts and consumables in the form of soft copy (e.g.: Microsoft EXCEL). The listing must have the following details; <ul style="list-style-type: none"> • Part Number • Part Description 		
WARRANTY PLANNED PREVENTIVE MAINTENANCE:	Tenderers must carry-out a minimum of two times PPM per year during the equipment warranty, starting six months after the date of commissioning.		
TECHNICAL TRAINING:	Tenderers to conduct training to Biomedical Engineers and Technicians.		
ON-SITE:	<ul style="list-style-type: none"> • Training to be conducted locally, tenderers are required to : <ul style="list-style-type: none"> • Provide training materials, test equipment, demo equipment, etc. • Provide training to two groups of technical staffs. • Provide 2 days (minimum) of training for each group. • Training to be conducted at Ministry of Health 		