

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
4	BME/A2K/(952)/2022/HTD (A19)	<p><b>TO SERVICES AND CALIBRATE INTERACOUSTICS AUDIOLOGY EQUIPMENTS FOR AUDIOLOGY UNIT, MINISTRY OF HEALTH</b></p> <p><u>PLACE OF SUBMISSION:</u>            QUOTATION BOX (GROUND FLOOR)            MINISTRY OF HEALTH            COMMONWEALTH DRIVE            BANDAR SERI BEGAWAN, BB 3910            NEGARA BRUNEI DARUSSALAM</p>	23/11/2022	10/12/2022	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	<p>MUHD AMIRUL FAZLEEN BIN HAJI KHALIDIN</p> <p>JURUTERA PERUBATAN HAYAT</p> <p>TÉL: 2242424</p> <p>EXT 8693 (RIPASH)</p>

# LAMPIRAN - A1

RUJ: BME/A2K/(952)/2022/HTD (A19)

9/11/2022

**TO SERVICE AND CALIBRATE INTERACOUSTICS AUDIOLOGY EQUIPMENTS FOR AUDIOLOGY UNIT,  
MINISTRY OF HEALTH**

## **SCOPE OF WORK**

### **EQUIPMENT DETAILS (10 UNITS)**

BRAND: INTERACOUSTIC

(MODEL, BME NUMBER, SERIAL NUMBER, LOCATION)

1. Affinity 2.0, BME9944, SN:0912358, Audiology unit (RIPASH)
2. Affinity 2.0, BME25574, SN:0991320, Audiology unit (RIPASH)
3. Affinity 2.0, BME25573, SN:0991321, Audiology unit (SSBH)
4. Affinity compact, BME27752, SN:1009014, Child Development Centre (CDC Kiarong Building 3)
5. Equinox PC based Audiometer, BME18621, SN: 0922418, Audiology unit (RIPASH)
6. Equinox PC based Audiometer, BME18620, SN: 0920071, Audiology unit (RIPASH)
7. Titan, BME25994, SN:0911559, Audiology unit (RIPASH)
8. Titan, BME27203, SN:1007324, Child Development Centre (CDC Kiarong Building 3)
9. Eclipse, BME27202, SN:1006383, Audiology unit (SSBH)
10. Eclipse, BME27201, SN:1007765, Child Development Centre (CDC Kiarong Building 3)

### **SCOPE OF WORK:**

**Must include but not limited to the following:**

- 1) ON-SITE CALIBRATION DONE BY LOCAL CERTIFIED ENGINEER OR OVERSEA ENGINEER WITH FACTORY APPROVED CALIBRATOR
- 2) AIR CONDUCTION
- 3) CALIBRATION OF PURE TONE AIR CONDUCTION
- 4) CALIBRATION OF NARROW BAND MASKING NOISE
- 5) BONE CONDUCTOR CALIBRATION
- 6) CHECKPOINTS:-
  - 6.1) CONTROL DEVICES E.G KNOBS AND PUSH BUTTONS
  - 6.2) INDICATING DEVICES E.G LAMPS AND DISPLAYS CORDS
  - 6.3) CORDS
  - 6.4) PLUGS
  - 6.5) HEADBANDS
  - 6.6) EARPHONES CUSHION
  - 6.7) HOSES AND TUBES FOR IMPEDANCE PROBES
- 7) PROVIDE ALL CERTIFICATION OF CALIBRATION FOR USERS AND BME