

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
7	BME/A2K/(958)/2022/HTD (A19)	<p>ON-OFF PPM AND PADS REPLACEMENT FOR TEN (10) UNITS OF MINDRAY AUTOMATIC EXTERNAL DEFIBRILLATOR (AED) AT HOSPITAL RIPAS AND VARIOUS LOCATION IN MINISTRY OF HEALTH</p> <p><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	23/11/2022	10/12/2022	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	MUHD AMIRUL FAZLEEN BIN HAJI KHALIDIN JURUTERA PERUBATAN HAYAT TEL: 2242424 EXT 8693 (RIPASH)

LAMPIRAN - A1

RUJ: BME/A2K/(958)/2022/HTD (A19)

14/11/2022

ONE-OFF PPM AND PADS REPLACEMENT FOR TEN (10) UNITS OF MINDRAY AUTOMATIC EXTERNAL DEFIBRILLATOR (AED) AT HOSPITAL RIPAS AND VARIOUS LOCATION IN MINISTRY OF HEALTH

SCOPE OF WORK

EQUIPEMENT DETAILS:

Model: Beneheart C1 (7 units)

(BME Number, Serial Number, Location)

1. BME23780, SN: AF3-01000547, WCC lobby
2. BME23781, SN: AF3-01000551, Cardiology Clinic
3. BME23782, SN: AF3-01000558, Canteen
4. BME23783, SN: AF3-01000543, Specialist Building lobby
5. BME23774, SN: AF3-01000559, State Medical Store Madaras
6. BME23775, SN: AF3-01000544, Pharmacy Madaras
7. BME23776, SN: AF3-01000546, Virology Lab

Model: Beneheart D1 (3 units)

8. BME23772, SN: FQ-02047494, BME Jetty
9. BME23761, SN: FQ-02047495, EMAS
10. BME23766, SN: FQ-02047496, EMAS

SCOPE OF WORK:

Must include but not limited to:

1. Visual inspection
2. Cleaning
3. Functional test
4. Performance test
5. Battery inspection
6. Inclusive of pads replacement
7. Apply preventive maintenance sticker/tag after successful service.
8. Vendor(s) to submit maintenance checklist/details/scope specific for Mindray Automatic External Defibrillator (AED) along with other necessary documents to join this tender.