

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
1	(131) PSD/QTN/2022 – (SSBH)	<p>“SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, FIVE (5) UNITS OF RESUSCITARE RADIANT WARMER FOR OBSTETRIC & GYNAECOLOGY AND PAEDIATRICS, HOSPITAL SURI SERI BEGAWAN, KUALA BELAIT”</p> <p><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	24/11/2022	10/12/2022	\$5.00	HOSPITAL SURI SERI BEGAWAN, KUALA BELAIT.	MOHAMAD ALIKHAN BIN HAJI MOHD YAKOP TEL: 3335331 EXT 2101/2102/2120

“SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, FIVE (5) UNITS OF RESUSCITARE RADIANT WARMER FOR OBSTETRIC & GYNAECOLOGY AND PAEDIATRICS, HOSPITAL SURI SERI BEGAWAN, KUALA BELAIT”

NO	DESCRIPTION OF ITEM	QUANTITY	BRAND	MANUFACTURER AND ORIGIN	UNIT PRICE	TOTAL PRICE
1	RESUSCITARE RADIANT WARMER	5 UNITS				
DELIVERY PERIOD: Not more than 4 weeks upon confirmation		(Yes/ No) (If No, please specify)				
PRICE VALIDITY : <i>The quotation shall remain valid for 6 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have the written consent of the suppliers.</i>						

“SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, FIVE (5) UNITS OF RESUSCITARE RADIANT WARMER FOR OBSTETRIC & GYNAECOLOGY AND PAEDIATRICS, HOSPITAL SURI SERI BEGAWAN, KUALA BELAIT”

TERMS AND CONDITIONS			
a.	Tenderer must be registered with the Ministry of Health	<p><i>Acknowledgement:</i></p> <p><i>Company Ref. No.:</i></p> <p>I hereby certify the above quote to be correct.</p> <p>Signature:</p> <p>.....</p> <p>Name:</p> <p>.....</p> <p>Designation:</p> <p>.....</p> <p>Date :</p> <p>.....</p>	Company's Official Stamp
b.	Please fill in the QUOTATION FORM completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form <u>may</u> cause DISQUALIFICATION OF QUOTATION		
c.	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION		
d.	Delivery Period: NOT MORE THAN 90 DAYS		
e.	Please do not use TIPPEX for amendment		

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**“SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, FIVE (5) UNITS OF
RESUSCITARE RADIANT WARMER FOR OBSTETRIC & GYNAECOLOGY AND
PAEDIATRICS, HOSPITAL SURI SERI BEGAWAN, KUALA BELAIT”**

	USER REQUIREMENT
FEATURES	• Infant radiant warmer with electrical variable height adjustment
	• With T-piece neonatal resuscitation
	• Adjustable breath range: 18-60 bpm
	• Precision blender 21-100% O2
	• Radiant warmer with microprocessor controls and surveillance
	• Servo and manual temperature control
	• Pre-warm, servo, and manual modes
	• Height adjustment with foot pedal
	• Resuscitation module & pressure gauge
	• Tilting bassinets with a large mattress
	• Folding down bed panels
	• Oxygen regulator & yoke
	• Adjustable built-in suction: 0-150mmHg
	• Suction gauge and suction bottle
	• APGAR timer
	• X-ray cassette tray
• Left/right swiveling warmer head for x-ray access	
• Visual display & overheat alarms	
• Tank mounting for 2 'E' cylinders	
ACCESSORIES/ CONSUMABLES (PER UNIT)	• Storage drawer
	• 1pc Reusable skin probe/ unit
	• Disposable T-piece breathing circuit – 10pcs/ unit

	SECTION II - PROCUREMENT REQUIREMENT
MODEL AND BRAND	• State Model and Brand of Equipment.
COUNTRY OF ORIGIN	• State Country of Origin.
PRICE (CIF) B\$	• State Unit and Total Price Offered.
WHERE MARKETED	• Specify in which other countries the equipment marketed.
YEAR MANUFACTURED	• State year of equipment manufactured.
WARRANTY	• State number of years.
DELIVERY TIME	• Not more than two months.