BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
6	46/QTN/JPK/2022	SUPPLY AND DELIVERY INSTALLATION TESTING AND COMMISSIONING ONE (01) UNIT OF PATIENT TRANSFER TROLLEY FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH	07/12/2022	24/12/2022	\$5.00	JABATAN PERKHIDMATAN KESIHATAN, KEMENTERIAN KESIHATAN.	HAJI AMIR HISMA BIN MASRI JURUTERA PERUBATAN HAYAT
		PLACE OF SUBMISSION: QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM					

	USER REQUIREMENT	
STANDARD FEATURES:		
	 Height range, cm: approximately 55 – 85 	
	Hydraulic foot pedals for height adjustment	
	Trendelenburg/ Reverse trendelenburg	
	 Trendelenburg: approximately -15° 	
	 Reverse trendelenburg: approximately +17° 	
	 Powered gas assisted backrest: 0 to +85° 	
	Radio translucent mattress platform	
	Central wheel braking system	
	200 mm anti-static castors with directional lock	
	 5th wheel steering system 	
	Integral fold down side rails	
	Dual pull/ push handles	
	Anti-bacterial framework	
	Easy to clean base	
	 Oxygen cylinder storage 	
	IV pole with 4 hooks with storage	
	Storage for patients' belongings	
	Inclusive of mattress	

	SECTION II - PROCUREMENT REQUIREMENT		
MODEL AND BRAND	State Model and Brand of Equipment.		
COUNTRY OF ORIGIN	State Country of Origin.		
PRICE (CIF) B\$	State Unit and Total Price Offered.		
WHERE MARKETED	Specify in which other countries the equipment marketed.		
YEAR MANUFACTURED	State year of equipment manufactured.		
WARRANTY	State number of years.		
DELIVERY TIME	Not more than two months.		
PRICE VALIDITY	State number of weeks.		

HEALTH CENTRE	DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH		
	TECHNICAL SPECIFICATIONS/REQUIREMENTS		
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	State equipment ambient operating temperature range.		
INTERNATIONAL SAFETY STANDARD:	 The equipment must comply to either one of the five international safety standard, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese JIS Standard. 		
TECHNICAL SUPPORT:	 State number of locally stationed engineers/technicians and the nearest overseas support. 		
DIMENSIONS:	State overall dimensions of equipment.		
WEIGHT:	State net weight of the equipment.		
EQUIPMENT WHOLE LIFE TIME SUPPORT:	 Supplier must guarantee spare parts are available at least for 10 years after the installation of the equipment. 		
BROCHURE:	 Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be- used (FORM B)" for reference. 		
USER AND SERVICE MANUALS:	 Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning. 		
SPARE-PARTS AND CONSUMABLES LISTINGS:	 Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description. 		
WARRANTY PLANNED PREVENTIVE MAINTENANCE:	 Minimum of two times PPM per year during the equipment warranty, six months after the date of commissioning and at the end of warranty period. 		
TRAINING:	 Tenderers are required to conduct technical training to Biomedical Engineers and Technicians. 		

	USER REQUIREMENTS	YES	NO	(IF 'YES') Please state catalogue/ brochure reference no.
STANDARD FEATURES:				
	Height range, cm: approximately 55 – 85			
	Hydraulic foot pedals for height adjustment			
	Trendelenburg/ Reverse trendelenburg			
	Trendelenburg: approximately -15°			
	 Reverse trendelenburg: approximately +17° 			
	Powered gas assisted backrest: 0 to +85°			
	Radio translucent mattress platform			
	Central wheel braking system			
	200 mm anti-static castors with directional lock			
	5 th wheel steering system			
	Integral fold down side rails			
	Dual pull/ push handles			
	Anti-bacterial framework			TO THE PERSON NAMED IN COLUMN
	Easy to clean base			
	Oxygen cylinder storage			
	IV pole with 4 hooks with storage			
	Storage for patients' belongings			
	Inclusive of mattress			

	SECTION II - PROCUREMENT REQUIREMENT	
BRAND AND MODEL OF EQUIPMENT:		
COUNTRY OF ORIGIN:		
PRICE (CIF) B\$	UNIT PRICE (B\$ CIF):	TOTAL PRICE (B\$ CIF):
WHERE MARKETED:		
YEAR MANUFACTURED:		
WARRANTY:		
DELIVERY TIME:		
PRICE VALIDITY:		

	SECTION III - TECHNICAL REQUIREMENTS
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	
INTERNATIONAL SAFETY STANDARD:	
TECHNICAL SUPPORT:	
DIMENSIONS (mm):	
WEIGHT (in kg):	
EQUIPMENT WHOLE LIFE TIME SUPPORT:	Number of years, spare parts are available after the installation of the equipment: years

	OBILL SENGRORONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES	YES	NO
BROCHURE:	Detailed brochure submitted?		
USER MANUALS:	Tenderers must supply during commissioning, three sets of user manuals, one set must be in the form of soft copy. (PDF Format		
SERVICE MANUALS:	Tenderers must supply during commissioning, three sets of service manuals, one set must be in the form of soft copy. (PDF Format).		
SPARE-PARTS AND CONSUMABLES LISTINGS:	Tenderers must supply during commissioning, a comprehensive list of equipment spare-parts and consumables in the form of soft copy (e.g.: Microsoft EXCEL). The listing must have the following details; Part Number Part Description		
WARRANTY PLANNED PREVENTIVE MAINTENANCE:	Tenderers must carry-out a minimum of two times PPM per year during the equipment warranty, starting six months after the date of commissioning.		
TRAINING:	Tenderers are required to conduct technical training to Biomedical Engineers and Technicians		