

LAMPIRAN 1

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
1	(04) PSD/QTN/2023 – (SSBH)	<p style="text-align: center;">“TO SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWENTY-FIVE (25) UNITS BLOOD/SOLUTION WARMING UNITS FOR SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT (CLUSTERING)”</p> <p style="text-align: center;"><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	25/01/2023	11/02/2023	\$5.00	HOSPITAL SURI SERI BEGAWAN, KUALA BELAIT.	DANIL BIN OSMAN PEGAWAI JURURAWAT TEL: 3335331 ext. 3250/ 3169

**“TO SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWENTY-FIVE (25) UNITS
BLOOD/SOLUTION WARMING UNITS FOR SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT (CLUSTERING)”**

NO	DESCRIPTION OF ITEM	QUANTITY	BRAND	MANUFACTURER AND ORIGIN	UNIT PRICE	TOTAL PRICE
1.	BLOOD WARMER	25 UNITS				
DELIVERY PERIOD: Not more than 4 weeks upon confirmation		(Yes/ No) (If No, please specify)				
PRICE VALIDITY : <i>The quotation shall remain valid for 6 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have the written consent of the suppliers.</i>						

**“TO SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWENTY-FIVE (25) UNITS
BLOOD/SOLUTION WARMING UNITS FOR SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT (CLUSTERING)”**

TERMS AND CONDITIONS			
a.	Tenderer must be registered with the Ministry of Health	<p><i>Acknowledgement:</i></p> <p><i>Company Ref. No.:</i></p> <p>I hereby certify the above quote to be correct.</p> <p><i>Signature:</i></p> <p>.....</p> <p><i>Name:</i></p> <p>.....</p> <p><i>Designation:</i></p> <p>.....</p> <p><i>Date:</i></p> <p>.....</p>	Company's Official Stamp
b.	Please fill in the QUOTATION FORM completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form <u>may</u> cause DISQUALIFICATION OF QUOTATION		
c.	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION		
d.	Delivery Period: NOT MORE THAN 90 DAYS		
e.	Please do not use TIPPEX for amendment		

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWENTY-FIVE (25) UNITS BLOOD/SOLUTION WARMING UNITS FOR SURI SERI BEGAWAN HOSPITAL

	USER REQUIREMENT
	<ul style="list-style-type: none"> Intended to be used during various clinical procedures to raise the temperature of refrigerated and room-temperature fluids as they are infused into the patient
STANDARD FEATURES:	<ul style="list-style-type: none"> Max temperature for Heat Exchanger should be 39 degrees Celsius
	<ul style="list-style-type: none"> Warming up time should be no more than 3 minutes
	<ul style="list-style-type: none"> Maximum flow at which blood can be heated from refrigerated state to should near body temperature should be no less than 15 mL/min
	<ul style="list-style-type: none"> High temperature cutoff at which blood is to be heated should be no more than 43 degrees Celsius
	<ul style="list-style-type: none"> Audio/Visual Alarms that should be included but not limited to: Overtemperature
	<ul style="list-style-type: none"> Must be able to utilize mains power supply of 220-240 V AC, 50-60 Hz
	<ul style="list-style-type: none"> Any accessories necessary for the daily operation of the device must be provided e.g. IV Pole attachment