

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
5	BME/A2K(989)/2023/HTD(A12)	<p>ONE-OFF PLANNED PREVENTIVE MAINTENANCE FOR INSTRAMED CARDIOMAX DEFIBRILLATORS IN MINISTRY OF HEALTH</p> <p><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	02/02/2023	25/02/2023	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	HAJAH SYAZWANI AZAH BINTI HAJI ABU BAKAR JURUTERA PERUBATAN HAYAT KANAN TEL: 2381740 ext. 7561 (HTD)

LAMPIRAN - A1

RUJ: BME/A2K/(989)/2023/HTD (A12)

28/1/2023

ONE-OFF PLANNED PREVENTIVE MAINTENANCE FOR INSTRAMED CARDIOMAX DEFIBRILLATORS IN MINISTRY OF HEALTH

SCOPE OF WORK

To conduct one-off PPM for the following Instramed Cardiomax defibrillators:-

- 1) BME18606, S/N: 30514 CM 4839 at Qashr Al Meezan Clinic
- 2) BME8353, S/N: 11113 CM 3238 at PIHM Hospital
- 3) BME8354, S/N: 11115 CM 3240 at PIHM Hospital
- 4) BME18121, S/N: 30512 CM 4837 at PIHM Hospital
- 5) BME14052, S/N: 20617 CM 3817 at RIPAS Hospital
- 6) BME12905, S/N: 30509 CM 4834 at RIPAS Hospital
- 7) BME12906, S/N: 30510 CM 4835 at RIPAS Hospital
- 8) BME8703, S/N: 11106 CM 3231 at SSB Hospital
- 9) BME8698, S/N: 11107 CM 3232 at SSB Hospital
- 10) BME8707, S/N: 11108 CM 3233 at SSB Hospital
- 11) BME8705, S/N: 11109 CM 3234 at SSB Hospital
- 12) BME8706, S/N: 11110 CM 3235 at SSB Hospital
- 13) BME8701, S/N: 11111 CM 3236 at SSB Hospital
- 14) BME8702, S/N: 11112 CM 3237 at SSB Hospital
- 15) BME8704, S/N: 11114 CM 3239 at SSB Hospital
- 16) BME8700, S/N: 11116 CM 3241 at SSB Hospital
- 17) BME8699, S/N: 11105 CM 3230 at SSB Hospital
- 18) BME9769, S/N: 30108 CM 4558 at SSB Hospital
- 19) BME9770, S/N: 30109 CM 4559 at SSB Hospital
- 20) BME9771, S/N: 30110 CM 4560 at SSB Hospital
- 21) BME9772, S/N: 30111 CM 4561 at SSB Hospital
- 22) BME9773, S/N: 30112 CM 4562 at SSB Hospital

Scope of work:

- i) Electrical Safety Test
- ii) Defibrillator Energy Output Test (Adult and Pediatric)
- iii) External Pacemaker BPM and Current calibration test
- iv) General Functionality Test:
 - a) SpO2
 - b) NIBP
 - c) Printer
- v) Check battery performance
- vi) Documentation of all test must be provided, including labelling with PPM sticker on equipment.